

Delivery/Customer information

Indole positive

Indole positive

Indole negative

Indole negative

-

CERTIFICATE OF ANALYSIS

PRODUCT	CM1046B BRILLIANCE E COLI/COLIFORM SEL MED 500g			202	Date Printed 2023.03.16 Delivery No.		
LOT NUMBER	3631339		Customer				
EXPIRY DATE 2026.03.31 DATE OF 2023.02.21				Cus	Customer Order Number		
Physical Characteristics		Results		Specifi	Specification		
Appearance Colour on reconstitution pH (25°C) Clarity		Straw pow	Straw powder Straw 2 6.9		Straw powder Straw 1-2 6.5 - 6.9		
		6.9					
		Clear		Clear			
Microbiological Pe	rformance	Control	Test	Recovery Test			
		cfu	cfu	%	Description		
Aerobic incubation	n at 37°C for 24 hours						
Escherichia coli ATCC®25922		33	42	127	Purple cols		
Escherichia coli ATCC®35218		46	49	107	Purple cols		
Klebsiella pneumoniae ATCC®29665		59	72	122	Pink cols		
Klebsiella pneumoniae ATCC®13883		55	58	105	Pink cols		
		Control	Test	Log(10) Reduction			
~ • •		cfu	cfu	_			
Staphylococcus aureus ATCC®25923		8.8E+04	0.0E+00	5	No growth		
Enterococcus faecalis ATCC®29212		4.4E+04	0.0E + 00	4	No growth		

Escherichia coli ATCC®25922 Escherichia coli ATCC®35218 Klebsiella pneumoniae ATCC®29665 Klebsiella pneumoniae ATCC®13883

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu). Negative strains are inhibited or shall produce at least a 2 log (10) reduction when compared to the control medium.

>1E+04

>1E+04

>1E+04

>1E+04

Refer to product specification for full details.





CERTIFICATE OF ANALYSIS

PRODUCT	CM1046B	
	BRILLIANCE E COLI/COLIFORM SEL MED 500)g
LOT NUMBER	3631339	

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release. Lot Accepted. 2023.03.02

Alarhen

.....

Carissa Courtney Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection. NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

