



# CERTIFICATE OF ANALYSIS

**PRODUCT** CM0961B  
BAIRD-PARKER AGAR BASE (modified) 500g

**LOT NUMBER** 3602075

**EXPIRY DATE** 2026.01.31

**DATE OF MANUFACTURE** 2022.12.20

**Delivery/Customer information**

Date Printed

2023.02.28

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 2	Straw 2
pH (25°C)	7.1	7.0 - 7.4
Clarity	Clear	Clear

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	-------------	----------	-----------------	-------------

**Tested with the addition of RPF Supplement SR0122****Aerobic incubation at 37 ± 2°C for 48 ± 2 hours**

*Staphylococcus aureus* ATCC®9144 53 62 117 Grey/black cols, coagulase zones

**Inoculation using diminishing sweep technique**

*Proteus mirabilis* ATCC®29906 >1E+04 - Straw/brown cols, no zones

**Testing has been performed in accordance with ISO11133:2014****Aerobic incubation at 37 ± 2°C for 48 ± 2 hours**

*S. aureus* ATCC®25923 WDCM00034 71 74 105 Grey/black cols, coagulase zones

*S. aureus* ATCC®6538 WDCM00032 65 66 102 Grey/black cols, coagulase zones

**Inoculation using diminishing sweep technique**

*S. saprophyticus* ATCC®15305 WDCM00159 3.2E+03 - Grey/black cols, no zones

*S. epidermidis* ATCC®12228 WDCM00036 5.5E+03 - Grey/black cols, no zones

*E. coli* ATCC®25922 WDCM00013 >1E+04 0 No growth

*E. coli* ATCC®8739 WDCM00012 >1E+04 0 No growth

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.  
Refer to product specification for full details.



Tested by the Quality Control Laboratory  
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England  
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096



## CERTIFICATE OF ANALYSIS

**PRODUCT** CM0961B  
**BAIRD-PARKER AGAR BASE (modified) 500g**  
**LOT NUMBER** 3602075

---

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2023.02.07

.....  
Carissa Courtney  
Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory  
**OXOID LIMITED**  
Wade Road, Basingstoke, Hampshire RG24 8PW. England  
oxid@thermofisher.com www.oxid.com