

Specification Sheet

Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by SAI Global Certificate Number QMS42099

Catalogue Number:	K/2000	Expiry Phrase:	Expires 2 yrs from packing date, see label.
--------------------------	--------	-----------------------	---

Description:	Aqualine complete 5		
Grade:		Application:	Karl Fischer volumetric reagent

Result Name	Specification	Units
Factor (mg/ml)	≥ 5.3 and ≤ 5.6	mg/ml

Additional Information:	A single component (composite) with a water equivalence of 5 mg water per ml. The factor result and specification apply at the time of testing only. The factor of this product will decrease over time
--------------------------------	---



Paula Faulkes

Mrs Paula Faulkes
Quality Assurance Manager.

Wednesday 21st of June 2023

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above. If there are any questions regarding this certificate, please call Technical Support on +32 1456 5600