

Certificate of Conformity

G-LAB Catalogue Number	9015D2-14G
Description	Petri dishes, PS, with 3 vents, 90 mm. Diameter, 15mm. Height, 14g., Sterile
Lot Number	20240812
Date of Manufacture	2024-08-08
Date of Sterilization	2024-08-20
Expiry Date	2026-08-11
Country of Origin	Manufactured in China
Date of Issue	2024-09-13

Quality System Compliance

Products are manufactured under the ISO 9001:2015 standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	Polystyrene (PS), Color : Transparent.
Sterilization	The products have the company of EO sterilizer, sterilizing the lowest temperature of 50°C, the highest temperature of 55°C, the dosage of 16kg, sterilization time 8 hours, testing qualified.
ATP Assay	Not Applicable.
DNase & RNase Free	Not Applicable.